

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 4057 Accredited to ISO/IEC 17025:2005	ISO-Tech Calibration Services Ltd	
	Issue No: 001 Issue date: 07 January 2008	
	Unit 3 Riverside Park East Service Way Raynesway Derby DE21 7RW	Contact: Mr P Avery Tel: +44 (0)1332 664245 Fax: +44 (0)1332 668399 E-Mail: Sales@iso96.demon.co.uk Website: www.iso-tech.org
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers premises	Dimensional	A



4057
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

ISO-Tech Calibration Services Ltd
Issue No: 001 Issue date: 07 January 2008

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MEASURING INSTRUMENTS AND MACHINES Performance verification of co- ordinate measuring machines	As ISO 10360-2 Up to 1500 mm (longest diagonal using length bars)	(1.5+1.4 per metre) μm See Note 1	NOTE 1. All calibrations must be carried out in accordance with procedures agreed by UKAS.	A

END